



Rev 6: rework burn-in power connector;
 add fuse between connector and regulator;
 add 5 components at pin 142 - VGN line of
 AMT 3 with signal return at pin 143 (VSS-TDC);
 add Weizmann Institute to silkscreen; add
 schottky diode and resistor to reset line, remove
 1N914.

MDT446			
File	MDT446		
Doc ID	Document Number	<Doc>	Rev 6
Date	Wednesday, April 23, 2003	Sheet	1 of 1